

Data Table

	Deviation
Frequency (Hz)	(%)
2	-2.3
3	1.9
4	3.6
5	3.8
6	4.1
4 5 6 7	4.4
8	4.4
9	4.2
10	4.4
20	2.7
30	1.7
50	0.5
75	0.1
100	0.0
200	-0.7
300	-1.3

Notes

Results relate only to the items calibrated.

This certificate may not be reproduced except in full, without written permission. Traceable to NIST (Project Number 822/271196) and PTB (Project Number 5399). This calibration was performed with TMS Portable Vibration Calibrator: S/N: 30220, Firmware version 3.3.4. Back-to-Back Comparison per ISO 16063-21

User Notes Utilized PCB Piezotronics model 480C02 to power velomitor CT

Customer:

Unit Condition As Found: In tolerance As Left: In tolerance

Technician: Algaissia

Approval: Algaissia

Approval Information

Lab Conditions Temperature: Humidity:

°C %

Calibration Date: 28-Apr-17 Calibration Time: 10:38 Due Date: